

# UNITED STATES PATENT AND TRADEMARK OFFICE



UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.		
09/890,778	01/24/2002	Thorsten Siess	IMPEL .57972	7976		
27629 75	90 02/17/2004		EXAM	EXAMINER		
	PATTON LEE & UTE ATE, SUITE 1550	BAXTER, J	BAXTER, JESSICA R			
LONG BEACH			ART UNIT	ART UNIT PAPER NUMBER		
	•		3731	71		

DATE MAILED: 02/17/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

	Application No.	Applicant(s)	1
	09/890,778	SIESS ET AL.	ON
Office Action Summary	Examiner	Art Unit	
	Jessica R Baxter	3731	
The MAILING DATE of this communicated for Reply	ation appears on the cover sheet with	h the correspondence ad	dress
A SHORTENED STATUTORY PERIOD FOR THE MAILING DATE OF THIS COMMUNIC.  - Extensions of time may be available under the provisions of after SIX (6) MONTHS from the mailing date of this commun.  - If the period for reply specified above, the maximum statu.  - Failure to reply within the set or extended period for reply will Any reply received by the Office later than three months after earned patent term adjustment. See 37 CFR 1.704(b).	ATION. 37 CFR 1.136(a). In no event, however, may a replication. days, a reply within the statutory minimum of thirty tory period will apply and will expire SIX (6) MONT II, by statute, cause the application to become ABA	ply be timely filed (30) days will be considered timely HS from the mailing date of this co	y. ommunication.
Status			
1) Responsive to communication(s) filed	on 08 December 2003.		
, —	o) ☐ This action is non-final.		
3) Since this application is in condition for	<i>'</i> —	rs, prosecution as to the	e merits is
closed in accordance with the practice			
Disposition of Claims			
4) ⊠ Claim(s) 8-12,18 and 19 is/are pendin 4a) Of the above claim(s) is/are 5) □ Claim(s) is/are allowed. 6) ⊠ Claim(s) 8-12, 18 and 19 is/are rejected. 7) □ Claim(s) is/are objected to. 8) □ Claim(s) are subject to restriction	withdrawn from consideration.		
Application Papers			
9) The specification is objected to by the	Examiner.		
10) The drawing(s) filed on is/are: a		y the Examiner.	
Applicant may not request that any objecti	ion to the drawing(s) be held in abeyand	e. See 37 CFR 1.85(a).	
Replacement drawing sheet(s) including the	he correction is required if the drawing(s	s) is objected to. See 37 CF	FR 1.121(d).
11)☐ The oath or declaration is objected to t	by the Examiner. Note the attached	Office Action or form PT	ГО-152.
Priority under 35 U.S.C. § 119			
	ocuments have been received. ocuments have been received in Ap f the priority documents have been r al Bureau (PCT Rule 17.2(a)).	oplication No received in this National	Stage
Attachment(s)			
1) Notice of References Cited (PTO-892)		ummary (PTO-413)	
<ul> <li>2) Notice of Draftsperson's Patent Drawing Review (PTO3) Information Disclosure Statement(s) (PTO-1449 or PPaper No(s)/Mail Date</li> </ul>		)/Mail Date formal Patent Application (PTC 	O-152)

### **DETAILED ACTION**

## Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

- (e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.
- 2. Claims 8, 9 and 19 are rejected under 35 U.S.C. 102(e) as being anticipated by U.S. Patent No. 6,248,091 to Voelker.

Voelker discloses a micro axial pump (pump 16) fastened to a catheter (catheter 10) and having a tubular pump portion comprising a pump ring (housing 17) and a cannula connected (distal end of support 13) therewith; and a dilating device (balloon 14) surrounding the pump portion wherein said microaxial pump comprises a tubular implantable drive portion that is directly coupled to said pump portion (shaft 23).

Regarding the limitation "wherein said dilating device is configured for breaking up a stenosis of a cardiac valve", Voelker's device is configured to dilate a stenosed valve, this is inherently break up the stenosis by forcing open the valve (see U.S. Patent NO. 4,808,165 Column 1 lines 28-33).

# Claim Rejections - 35 USC § 103

- 3. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:
  - (a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the

Application/Control Number: 09/890,778

Art Unit: 3731

Page 3

subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

4. Claims 10, 11, 12 and 18 are rejected under 35 U.S.C. 103(a) as being unpatentable over Voelker '091 in view of U.S. Patent No. 4,808,165 to Carr.

Voelker discloses the claimed invention except for the particular pressure to which the balloon is inflated. Carr teaches that a balloon used dilating stenosed cardiac valves is inflated to at least 1.0 bar. These high pressures are necessary to open the stenosed valve (Column 1 lines 34-40). It would have been obvious to one having ordinary skill in the art at the time the invention was made to provide the device of Voelker with the high inflation pressure of Carr in order to allow the device of Carr to dilate the stenosed valve.

### Response to Arguments

- 5. Applicant's arguments filed December 8, 2003 have been fully considered but they are not persuasive.
- 6. Applicant argues that Voelker '091 does not disclose a catheter adapted for breaking up a stenosed valve. In Column 1 lines 34-40, Voelker states that the balloon catheter dilates a stenosed valve (Column 1 lines 31-43 and Column 2 lines 6-10), inherently breaking up the stenosis (see U.S. Patent NO. 4,808,165 Column 1 lines 28-33). Applicant argues that Voelker's drive portion is not implantable. The drive *portion* (shaft 23) is placed into the blood vessel. The drive portion is thus implanted into the body. Therefore, the rejection over Voelker '091 is deemed proper.

#### Conclusion

7. THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Jessica R Baxter whose telephone number is 703-305-4069. The examiner can normally be reached on M-F 8:30AM - 5:00PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Michael Milano can be reached on 703-308-2496. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Jessica R Baxter Examiner Art Unit 3731

MICHAEL J. MILANO SUPERVISORY PATENT EXAMINER TECHNOLOGY CENTER 3700

MB Irb